

Монитор кислорода и углекислого газа SUB Mk II F

Руководство по эксплуатации

По вопросам продаж и поддержки обращайтесь:

Алматы (727)345-47-04
Ангарск (3955)60-70-56
Архангельск (8182)63-90-72
Астрахань (8512)99-46-04
Барнаул (3852)73-04-60
Белгород (4722)40-23-64
Благовещенск (4162)22-76-07
Брянск (4832)59-03-52
Владивосток (423)249-28-31
Владикавказ (8672)28-90-48
Владимир (4922)49-43-18
Волгоград (844)278-03-48
Вологда (8172)26-41-59
Воронеж (473)204-51-73
Екатеринбург (343)384-55-89

Иваново (4932)77-34-06
Ижевск (3412)26-03-58
Иркутск (395)279-98-46
Казань (843)206-01-48
Калининград (4012)72-03-81
Калуга (4842)92-23-67
Кемерово (3842)65-04-62
Киров (8332)68-02-04
Коломна (4966)23-41-49
Кострома (4942)77-07-48
Краснодар (861)203-40-90
Красноярск (391)204-63-61
Курск (4712)77-13-04
Курган (3522)50-90-47
Липецк (4742)52-20-81

Магнитогорск (3519)55-03-13
Москва (495)268-04-70
Мурманск (8152)59-64-93
Набережные Челны (8552)20-53-41
Нижний Новгород (831)429-08-12
Новокузнецк (3843)20-46-81
Ноябрьск (3496)41-32-12
Новосибирск (383)227-86-73
Омск (3812)21-46-40
Орел (4862)44-53-42
Оренбург (3532)37-68-04
Пенза (8412)22-31-16
Петрозаводск (8142)55-98-37
Псков (8112)59-10-37
Пермь (342)205-81-47

Ростов-на-Дону (863)308-18-15
Рязань (4912)46-61-64
Самара (846)206-03-16
Санкт-Петербург (812)309-46-40
Саратов (845)249-38-78
Севастополь (8692)22-31-93
Саранск (8342)22-96-24
Симферополь (3652)67-13-56
Смоленск (4812)29-41-54
Сочи (862)225-72-31
Ставрополь (8652)20-65-13
Сургут (3462)77-98-35
Сыктывкар (8212)25-95-17
Тамбов (4752)50-40-97
Тверь (4822)63-31-35

Тольятти (8482)63-91-07
Томск (3822)98-41-53
Тула (4872)33-79-87
Тюмень (3452)66-21-18
Ульяновск (8422)24-23-59
Улан-Удэ (3012)59-97-51
Уфа (347)229-48-12
Хабаровск (4212)92-98-04
Чебоксары (8352)28-53-07
Челябинск (351)202-03-61
Череповец (8202)49-02-64
Чита (3022)38-34-83
Якутск (4112)23-90-97
Ярославль (4852)69-52-93

Россия +7(495)268-04-70

Казахстан +(727)345-47-04

Беларусь +(375)257-127-884

Узбекистан +998(71)205-18-59

Киргизия +996(312)96-26-47

эл.почта: axq@nt-rt.ru || сайт: <https://analox.nt-rt.ru/>

LIST OF CONTENTS

1	PACKAGING CONTENTS CHECK	3
2	INTRODUCTION	4
3	INSTALLATION	6
3.1	MAIN PANEL MECHANICAL INSTALLATION	6
3.2	REMOTE SENSOR MECHANICAL INSTALLATION	8
3.3	SYSTEM ELECTRICAL INSTALLATION	9
4	OPERATION	15
4.1	SWITCHING ON AND OFF	16
4.2	MAIN DISPLAY SCREEN	17
4.3	ALARM SYSTEM	18
4.4	OXYGEN INJECTION SYSTEM.....	23
4.5	FUNCTION KEYS	25
4.6	ALARM SETPOINT MENU.....	26
4.7	CALIBRATION MENU	29
4.8	INTERFACE CALIBRATION MENU	36
4.9	SETUP MENU	38
4.10	OXYGEN INJECTION MENU	39
5	DATALOGGING	40
5.1	RS232 DATA LOGGING CONNECTION.....	40
5.2	USE OF WINDOWS HYPERTERMINAL	42
5.3	RETRIEVE DATA: GET COMMAND.....	43
5.4	DELETE DATA: DEL COMMAND	43
5.5	USE OF RETRIEVED DATA	44
6	CALIBRATION.....	44
6.1	SEMI AUTOMATIC GAS CALIBRATION	45
6.2	PRESSURE SENSOR CALIBRATION	46
6.3	OXYGEN SENSOR CALIBRATION	46
6.4	CARBON DIOXIDE SENSOR CALIBRATION.....	47
6.5	TEMPERATURE SENSOR CALIBRATION	47
6.6	HUMIDITY SENSOR CALIBRATION	49
6.7	INTERNAL CALIBRATION	50
7	TROUBLESHOOTING.....	58
8	SAFETY WARNINGS.....	59
9	SPECIFICATION.....	60
10	DISPOSAL	49

1 PACKAGING CONTENTS CHECK

- a) Analox Sub Mk II F main unit.
- b) Analox Sub Mk II F REM1 Oxygen, Carbon Dioxide and Depth Remote Sensor Unit
- c) Optional Analox Sub Mk II REM2 Temperature and Humidity Remote Sensor Unit
- d) Optional Analox Sub Mk II REM3 Oxygen Injection Controller Unit with control cable
- e) Optional Mains Cable (AC powered instruments only)
- f) Connecting cables to connect Main Unit to REM1 and optional REM2/REM3 units
- g) Calibration adaptors and tubing
- h) Optional Datalogging cable
- i) User Manual
- j) Test Certificate

***Please refer to the Safety Warning information in Section 8
and to the Caution notice in Section 3.3.***



2 INTRODUCTION

The Analox Sub Mk II F is a combined oxygen, carbon dioxide (CO₂) and pressure monitor that automatically compensates the CO₂ data for pressure effects. Optionally, temperature and humidity may also be measured if fitted at the time of manufacture. An optional controller module may also be added to provide oxygen injection to maintain oxygen levels at preset levels.

Parameter	Sensing technology	Sensor Location
Oxygen	Electrochemical cell	REM1
Carbon Dioxide	Infra-red absorption	REM1
Pressure	Strain gauge	REM1
Temperature	Platinum Resistor	REM2
Humidity	Capacitive	REM2
Oxygen (Control Sensor)	Electrochemical cell	REM3

The complete system consists of:

- a) The SUB Mk II F main unit intended for installation in control rooms (special versions are available if it is required to mount the Main Unit in the hyperbaric environment).
- b) The REM1 and optional REM2 and/or REM3 intended for installation in hyperbaric environments.

A large graphic display on the main unit shows the value of each measured parameter and the remote unit provides local displays of the monitored parameters.

The main panel can provide audio and visual alarms for the measured gas parameters. The alarm setpoints may be easily adjusted using the pushbuttons on the main panel.

The overall system is powered by a single supply as indicated on the rear of the main unit. The remote unit connects to the main panel, from which it obtains power.

The sensors are housed in splashproof enclosures that are vented to prevent collapse in hyperbaric environments. Gas levels are monitored by diffusion across waterproof membranes built into the REM1 (and REM3 when fitted) unit.



3 INSTALLATION

3.1 MAIN PANEL MECHANICAL INSTALLATION

The Main Panel unit is designed to be panel mounted. The panel must first be prepared to accept the instrument by making the cutout as shown below.




Figure 1: Main Panel - Panel Cutout

The shaded area represents the aperture into which the main panel will fit. The 4 mounting holes are to be drilled 10mm diameter. Four M5 captive nuts and four M5 screws are supplied with the instrument to mount it to the panel.

If the unit is likely to be subjected to high levels of vibration, it is recommended to construct a supporting bar or shelf towards the rear of the unit. The two figures below show a suitable means of providing this support. The absolute position of the shelf should be anywhere within 50mm of the rear of the instrument.

Note that additional space will be required behind the panel for the mating connectors.




Figure 2: MAIN PANEL REAR SHELF – VIEW FROM ABOVE




Figure 3: MAIN PANEL REAR SHELF – VIEW FROM REAR

3.2 REMOTE SENSOR MECHANICAL INSTALLATION

The REM1 and REM2 are designed to be wall mounted using the mounting lugs fitted as shown below. The enclosures are fitted with a vent to prevent collapse in hyperbaric conditions. Ensure that the vent is not obstructed

Ensure that the remote sensor units are positioned sensibly with regard to the gases being monitored. In the absence of air circulation, the gas sensors should be mounted at breathing height. They should be mounted against a near vertical surface

If they were mounted flat, there would be a possibility that water will collect in the sensor inlets and block the diffusion of gas into and out of the sensor.




Figure 4: REM MOUNTING DIMENSIONS

3.3 SYSTEM ELECTRICAL INSTALLATION

Caution

It is advised to ensure that the system is always switched off whenever making or modifying the electrical connections to the system. Mains powered instruments should be unplugged from the mains for personal safety. All instruments (AC and DC powered) should be powered down. There is a possibility of damaging the system by tampering with live connections.

The Main Panel unit requires an electrical supply. Mains powered versions are provided with a mains cable. DC powered versions have rear panel terminals to accept a DC supply, and an internal line filter. Terminals are provided on AC powered versions to also allow them to operate from DC supplies, but there is no internal line filter fitted.

An Earth terminal is also provided, although mains powered systems can be earthed through the mains cable. The earth is only used for screening purposes.

A 2 metre connecting lead is supplied which must connect between the Main Panel and the Remotes Sensors as shown below. The cable is supplied with a break in the middle, with the wires temporarily connected through a terminal connector. This allows the user to 'break' the cable for wiring through the hull of pressure chambers.

If there are no REM2 or REM3 sensors fitted, the cable connects directly into REM1. Additional cables supplied connect the REM1, REM2 and REM3 (when fitted) sensors together in a "daisy-chain" or style.

The cable screen may be required in certain installations to minimise interference with other systems (voice communications etc). The cable screen is connected to the earth of the Main Panel, and should be continued through to the remote sensors. Some special versions of the remote sensors are built into metal enclosures which may themselves be earthed. In such cases, it may sometimes help to disconnect the screen connection at the terminal block to avoid the presence of an earth loop.

The cable is a two pair, overall screened, flame retardant (BS4066 and IEC332 Part 3) cable. Each core is 7 strands of 0.254mm diameter wire (csa 0.35mm²). Overall cable diameter is nominally 4.2mm.







Figure 5: SYSTEM WIRING

3.3.1 MAIN PANEL ELECTRICAL CONNECTIONS



All electrical connections are situated on the rear of the main panel. Depending upon the configuration ordered by a customer, not all of the connections may be fitted. Also there may be no internal connections to some of the terminals. For instance, a unit without relay outputs will not have anything connected to the relay output terminals.

Rear Panel Terminals permit connection of the following (The provision of these terminals depends on the combination of options ordered)

3.3.1.1 AC Power Input (Optional)

The connector for an AC mains input is only fitted to AC powered instruments. A standard IEC type connector and cable is supplied with the instrument. A plug will be fitted to the cable appropriate to the country to which the instrument is first delivered (UK, US, European or Australian). If an alternative plug type is required, either contact Analox for assistance, or source the correct type locally.

3.3.1.2 DC Power Input

DC voltage in the range 12v to 24v DC may be applied to terminals 1 to 4. No connection is necessary for AC mains powered instruments. These terminals may still be used to temporarily power an AC instrument from DC, although note that there is no line filter fitted to the DC input in this configuration.

1	2	3	4
DC Power Supply Input			
+24V		0V	

Terminals 1 and 2 (positive input) are commoned together internally. Terminals 3 and 4 (zero volt input) are commoned together internally. This permits power to be looped into and out of the equipment where necessary.

3.3.1.3 Relay Outputs (Optional)

The system can provide up to six relay outputs. Relays 1 to 4 provide volt-free single pole changeover contacts, while Relays 5 and 6 provide volt-free single pole normally open contacts.

.(if it is essential to have the normally closed contact from either Relay 5 or Relay 6, this can be arranged by modifying internal wiring -consult the manufacturer for details) .

All of the connections are provided on terminals 15 through to 30 as shown below:

15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30
RELAY1			RELAY2			RELAY 3			RELAY 4			RELAY 5		RELAY 6	
NO	COM	NC	NO	COM	NC	NO	COM	NC	NO	COM	NC	NO	COM	NO	COM

Each relay is internally assigned to an alarm condition, or group of alarm conditions.

The following table shows the standard allocations, which may be altered via a configuration file within the system.



SYSTEM	RELAY1	RELAY2	RELAY 3	RELAY 4	RELAY 5	RELAY 6
REM 1	Oxygen Low	Oxygen Hi	CO2 Hi	CO2HiHi	Depth Low	Depth High
REM1+ REM2	Oxygen Low	Oxygen Hi	CO2 Hi	CO2HiHi	Temperature Lo or Hi	Humidity Lo or Hi

The relays are set as standard in 'Fail Safe' mode. With the instrument switched off, the relays are de-energised, and this is taken as the alarm state. Therefore when switched on, the relays will be energised whenever there is NO alarm present. Alternatively, the relays may be set in non 'Fail-Safe' mode, when the relays are energised in alarm.

In the standard configurations the following will apply

- a) an oxygen relay will also show an alarm if there is a calibration fault on the sensor, or if there is no communications with REM1.
- b) a CO2 relay will also show an alarm if there is a calibration fault on the sensor, or if there is no communications with REM1.
- c) a pressure/depth relay will also show an alarm if there is a calibration fault on the sensor, or if there is no communications with REM1.
- d) a temperature or humidity relay will also show an alarm if there is a calibration fault on the sensor, or if there is no communications with REM2.

Note the allocation of alarms to each relay output is configurable at the time of manufacture, as is the 'Fail Safe' mode. The settings supplied may differ to those shown above if previously arranged with the manufacturer.

3.3.1.4 Analogue Outputs (Optional)

The system provides up to five 4-20mA output channels. Voltage outputs are also available by request.

These provide active outputs (as opposed to passive). The 4-20mA signal is sourced by the Analox equipment. Only passive equipment should be connected to these outputs.

The default factory settings are to assign each of the outputs to each of the measured parameters as shown below. 4mA is set to represent the zero of the measured parameter, whilst 20mA is set to represent the full scale of the measured parameter.

As an example, for a 0-2000mBar pO2 channel, 4mA will represent 0mBar, and 20mA will represent 2000mBar.

5	6	7	8	9	10	11	12	13	14
CH1	CH2	CH3	CH4	CH5					



+	-	+	-	+	-	+	-	+	-
OXYGEN		CO2		DEPTH		TEMP		HUMIDITY	

The internal configuration file allows these settings to be altered if for instance a customer perhaps wants 4-20mA to represent say 100-500mBar pO₂. The internal configuration also allows for instance if two oxygen outputs were required etc.

Note the allocation of parameters to each channel is the default setting for a system fitted with REM1 and REM2 sensors. The settings supplied may differ to those shown above if previously arranged with the manufacturer.

3.3.1.5 RS232 Datalogging Port (Optional)

Please refer to Section 5 for details of this connection.

3.3.2 REM3 OXYGEN INJECTION INSTALLATION

The REM3 is supplied with a 2m cable which connects a volt-free contact within the REM3 into the customers control wiring for the injection solenoid as shown below.

For further assistance regarding solenoids, please contact Analox.




Figure 6: INJECTION CONNECTIONS

4 OPERATION

The Main Panel provides

- a) a 240x128 pixel liquid crystal display
- b) function key pushbuttons F1-F5 which align with key legends on the display
- c) a yellow alarm indicator lamp
- d) a red fault indicator lamp
- e) a contrast control for the display
- f) a brightness control for the display back light
- g) an Acknowledge pushbutton to acknowledge alarms and faults
- h) an audible sounder to annunciate alarms and faults
- i) a power switch controlling power to the instrument and to the remote sensor units
- j) a fuse F1 for the Main Panel
- k) a fuse F2 for the Remote Sensors




Figure 7: MAIN PANEL DETAIL

4.1 SWITCHING ON AND OFF

The Main Panel and the Remote Sensors are switched on by operating the Power switch on the Main panel. The switch has a locking device to prevent inadvertent operation. Gently pull the lever outwards and then down to turn the power on, and outwards and up to turn the power off. Releasing the lever re-applies the locking mechanism.

If the unit does not switch on, check the following:

- a) check that the external power supply is healthy.
- b) check that the supply fuse on the rear panel is not blown (mains powered instruments only).
- c) check that the front panel fuses are not blown.

Several events take place at switch on. These are:

- a) the audible alarm sounds for several seconds while the processor initialises. This ensures that in the event of a system crash via the on-board Watchdog controller, an operators attention will be attracted by the audible alarm. . (This can be disabled if necessary - consult Analox and request information sheet AS2F-046).
- b) both the Fault and Alarm indicators are illuminated to prove their operation
- c) all pixels on the display are turned on and then off as a test of the display
- d) a version number message is displayed on the screen as shown below. Note the version number on the screen may change as upgrades to the system are made.

Please wait while system initialises....

- e) the screen is then cleared and the main application screen appears . It is quite normal, especially for carbon dioxide, for the sensors to take a short time to warm up and give stable readings. Alarms related to the sensor readings are inhibited for the first 40 seconds of operation to give the sensors time to settle.
- f) If connected correctly, the remote units will briefly display '.8.8.8.8' on each of their displays at power on, followed by sensor readings after a brief delay.

Note that for several seconds after power on, the display screen is unable to be updated by the internal computer. As soon as the computer initialises, the screen is updated. The appearance of the screen during this period is random. It may flicker, but please wait for the system to power up completely before assuming this is a problem. Normal time is around 40 seconds. If it exceeds 1 minute, then something is wrong.

4.2 MAIN DISPLAY SCREEN

The Main Display screen is shown below

HH:MM:SS LOG OFF 1m 4%		RLY	
OXYGEN (mBar) 209 			

Figure 8: MAIN DISPLAY SCREEN

The screen is divided into several parts as follows:

- a) along the bottom of the screen, there are legends which align with the function keys.
- b) the left hand side of the screen shows sensor readings for REM1 . Note that the range and units of measurement may change for specific instruments.

- c) readings in alarm states are highlighted with a black background (reverse video). These readings correspond to those shown on the Remote Sensors.
- d) alongside the oxygen and carbon dioxide readouts, the present settings of the alarm setpoints are shown. < signifies a Low alarm, > a high alarm and >> a very high alarm.
- e) the time of day is displayed at the top of the display
- f) data logging status information is shown at the top of the display. This states whether the Logging is turned ON or OFF, the logging interval in either minutes or seconds, and the %disk capacity utilised for logging
- g) an activity indicator in the bottom right hand corner which shows that the processor is functioning.
- h) the right hand side of the screen shows sensor readings for REM2 (with high and low alarm setpoints) and REM3 (ON/OFF status, setpoint and control value). Note these are options and may not be fitted to a particular system.

4.3 ALARM SYSTEM

The primary alarm conditions monitored by the system are the oxygen, carbon dioxide, pressure (or depth), temperature and humidity alarms. These conditions are referred to as Alarms in the system.

The design of the system is such that in trying to gather the correct information, various faults that may occur can be detected. These conditions are referred to as Faults in the system.

Both Alarms and Faults cause the Audible Sounder to pulse.

An Alarm condition occurring causes the 'ALARM' Indicator on the Main Panel to flash.

A Fault condition occurring causes the 'FAULT' Indicator on the Main Panel to flash.

The Acknowledge switch on the Main Panel may be pressed to silence the audible sounder. It will also cause the Fault and/or Alarm Indicators to stop flashing. If the alarm or fault condition still exists, the Alarm or Fault indicator will remain turned on.

When the Alarm or Fault condition ceases to exist, the alarm or fault indicators will be turned off and then remain turned off.



The system is arranged with non-latching alarms. This means that if an alarm or fault condition occurs, but then ceases to exist, then the Audible sounder will silence and the indicators will turn off without the Acknowledge switch having been pressed.

The Calibration Error Faults are reported to the Remote Sensor so that a local indication is given that calibration is required.



4.3.1 LIST OF ALARM CONDITIONS

The Table below provides details of all of the alarm conditions annunciated by the system.

ALARM NAME	CONDITION CAUSING ALARM
Oxygen Low	Oxygen content is less than the low alarm setpoint
Oxygen High	Oxygen content is greater than the high alarm setpoint
Carbon Dioxide High	CO2 content is greater than the high alarm setpoint
Carbon Dioxide Very High	CO2 content is greater than the very high (HI2) alarm setpoint
Depth Low	Depth is less than the low alarm setpoint
Depth High	Depth is greater than the high alarm setpoint
Temperature Low	Temperature is less than the low alarm setpoint
Temperature High	Temperature is greater than the high alarm setpoint
Humidity Low	Humidity is less than the low alarm setpoint
Humidity High	Humidity is greater than the high alarm setpoint

Whenever any of these alarm conditions are recognised, the appropriate reading on the display is highlighted on a Black background (reverse video). The corresponding Remote Display will flash.

Hysteresis is applied to each of the alarm thresholds. The hysteresis band is pre-set to 2% of the alarm setpoint. Thus if a low oxygen alarm setpoint is set at 180mBar, the alarm will be raised when the oxygen level falls below 180mBar oxygen, and the alarm will be cleared when the oxygen level rises to 180mBar oxygen +2% (=183.6mBar oxygen).

Similarly if a high oxygen alarm setpoint is set at 230mBar, the alarm will be raised when the oxygen level rises above 230mBar, and the alarm will be cleared when the oxygen level falls to 230mBar oxygen -2% (=225.4mBar oxygen).

4.3.2 LIST OF FAULT CONDITIONS

The table below provides details of all of the fault conditions annunciated by the system.

FAULT NAME	CONDITION CAUSING FAULT	INDICATED BY	SUGGESTED REMEDIAL ACTION
REM1 Communication Timeout	Main Panel cannot communicate with the REM1 Sensor Unit	COMMS ERROR appears below REM1 heading, and sensor readings are blanked out on Main Panel	Ensure that REM1 sensor is connected and powered up. If it is, check the wiring is correctly made.
REM2 Communication Timeout	Main Panel cannot communicate with the REM2 Sensor Unit	COMMS ERROR appears below REM2 heading, and sensor readings are blanked out on Main Panel	Ensure that REM2 sensor is connected and powered up. If it is, check the wiring is correctly made.
REM3 Communication Timeout	Main Panel cannot communicate with the REM3 Sensor Unit	REM3COMMS appears in place of Cv in Injection display box	Ensure that REM3 sensor is connected and powered up. If it is, check the wiring is correctly made.
Interface Module Communication Timeout	Main Panel cannot communicate with its own internal Interface module (relay and Analog Outputs)	RLY appearing in the upper right hand corner of the display	Disconnect REM1m REM2 and REM3 from Main Panel. If fault still remains, line driver within Main Panel probably needs replacing
O2 Cal-H	The Main Panel has determined that the sensor reading is invalid when a High point calibration has been performed for a REM1 sensor (or for a REM3 oxygen sensor)	Cal-H appears in the O2 display	Recalibrate the appropriate sensor
Depth (Pressure) Cal-H		Cal-H appears in the Depth display	
Temperature Cal-H		Cal-H appears in the Temperature display	
Humidity Cal-H		Cal-H appears in the Humidity display	
O2 Cal-L	The Main Panel has determined that the sensor reading is invalid when a Low point calibration has	Cal-L appears in the O2 display	Recalibrate the appropriate sensor
Depth (Pressure) Cal-L		Cal-L appears in the Depth display	

FAULT NAME	CONDITION CAUSING FAULT	INDICATED BY	SUGGESTED REMEDIAL ACTION
Temperature Cal-L		Cal-L appears in the Temperature display	
Humidity Cal-L		Cal-L appears in the Humidity display	
CO2 Cal-S	The Main Panel has determined that the CO2 sensor reading is invalid when a Span calibration has been performed for a REM1 sensor	Cal-S appears in the CO2 display	Recalibrate the CO2 sensor
CO2 Cal-0	The Main Panel has determined that the CO2 sensor reading is invalid when a Zero calibration has been performed for a REM1 sensor	Cal-0 appears in the CO2 display	Recalibrate the CO2 sensor
CO2 Flt-S	The REM1 has detected a fault thought to be due to the Infra -red source	FltS appears in the REM1 CO2 display	Attempt recalibration of CO2 sensor, or contact factory
CO2-Flt-d	The REM1 has detected a fault thought to be due to the Infra -red detector	Fltd appears in the REM1 CO2 display	Attempt recalibration of CO2 sensor, or contact factory

4.4 OXYGEN INJECTION SYSTEM

The injection unit will power up with the main power switch on the SubMkIIIF front panel

A local switch on the sensor enclosure will Enable/Disable the Control action. When in the Disable Position, control action will be inhibited. When in Enable position, operation will be controlled by the Main Panel control. In the event of a fault, it is possible to force the REM3 into a controlling action by moving the switch from Enable to Disable and back to Enable.

The Injection module is designed to operate a solenoid valve to control the flow of oxygen to the area being monitored. If the measured oxygen value is greater than the injection setpoint, then no action is taken by the instrument.

The standard range of oxygen measurement is 0-2000mBar O₂. The setpoint can be set to any value in this range from the Main Panel.

If the measured value of oxygen falls below the Injection setpoint, then the solenoid valve is activated and oxygen is injected for a time period proportional to the difference between the setpoint and the measured value. On completion of the injection period, the module allows a Pre-set Mixing Period of x seconds (initially guess at 100) to permit mixing of the recently injected oxygen. The Mixing time is adjustable from the Main Panel.

Opening Times for the injection valve are settings in an EEROM based Lookup table are:

Error (Setpoint-Oxygen) mBar	Valve Opening Time (seconds)
<0	0
0	0.5
1	1.0
2	2.0
3	3.0
4	5.0
5	10.0
6	20.0
7	32.0
8	46.0
9	68.0
>=10	100.0



It should not be necessary to adjust these. The amount of oxygen injected can be preset by the pressure of the oxygen regulator. If these times do need to be changed, contact Analox for advice.

Mixing Time is expected to be in the nominal range of 30-120 secs. It is essentially determined by the size of the chamber and the positioning of injection points within the chamber.

After the Mixing period, a further check is made of the oxygen concentration to determine if the solenoid valve should be re-opened.

When the oxygen level is above the setpoint, no injection is required. The mixing period is therefore not required. The unit will wait for the oxygen level to fall to a value at which further injection is required and then immediately perform an injection cycle.

The LED indicator on the REM3 Injection unit remains on during a mixing period, and flashes during an injection period. It will remain OFF whilst injection is disabled.

4.5 FUNCTION KEYS

When first powered on, the Main Menu appears as shown below:



Press 'ALARM SETPT' to adjust alarm settings for any of the gas alarms

Press 'CALIB' to perform a calibration of any of the sensors

Press 'SETUP' to enter the SETUP menu.

Press 'O2 INJCT' to enter the Oxygen Injection menu (Not available unless REM3 is fitted)

Press 'LOG TOGGLE' to turn the Data Logging On or Off. Note it cannot be turned on if the storage memory is full.

(Note that a datalogging cable is only supplied with instruments for which datalogging is specified at the time of ordering).

4.6 ALARM SETPOINT MENU

From the Main Menu, pressing 'ALARM SETPT' brings up a new menu requesting the user to select the zone in which the alarm setpoints are to be altered:

Select REM1 setpoint

O2	CO2	DEPTH	MORE	MAIN
----	-----	-------	------	------

Pressing 'MORE' brings up a new menu for the REM2 sensor setpoints

Select REM2 setpoint

TEMP	HUMID		MORE	MAIN
------	-------	--	------	------

On either of these screens, pressing O2, CO2, Depth, Temp or Humid will bring up the corresponding menu for that sensor as shown below.

Select REM1 Oxygen setpoint

O2-LO	O2-HI		EXIT	MAIN
-------	-------	--	------	------

Select REM1 CO2 setpoint

CO2H1	CO2H2		EXIT	MAIN
-------	-------	--	------	------

Select REM1 Depth/Pres setpoint

PR-LO	PR-HI		EXIT	MAIN
-------	-------	--	------	------

Select REM1 Temperature setpoint

TMPLO	TMPHI		EXIT	MAIN
-------	-------	--	------	------

Select REM1 Humidity setpoint

HUMLO	HUMHI		EXIT	MAIN
-------	-------	--	------	------

On any of these screens, pressing EXIT will return to the previous menu, whilst pressing MAIN will return to the main menu.

Press 'O2-LO' to adjust the Oxygen low alarm setpoint. The present value of the alarm is displayed as shown below.

Press 'O2-HI' to adjust the Oxygen high alarm setpoint. The present value of the alarm is displayed as shown below.



Press 'CO2HI' to adjust the carbon dioxide high alarm setpoint. The present value of the alarm is displayed as shown below.

Press 'CO2HI2' to adjust the carbon dioxide very high alarm setpoint. The present value of the alarm is displayed as shown below.

Press 'PR-LO' to adjust the Pressure/Depth low alarm setpoint. The present value of the alarm is displayed as shown below.

Press 'PR-HI' to adjust the Pressure/Depth high alarm setpoint. The present value of the alarm is displayed as shown below.

REM1 O2LO 180 mBar

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM1 O2HI 230 mBar

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM1 CO2HI2 15.0 mBar

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM1 CO2HI1 5.0 mBar

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM1 PRLO 0.50 BarA

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM1 PRHI 6.00 BarA

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM2 TMPLO 10.0 °C

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM2 TMPHI 30.0 °C

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM2 HUMLO 30 %RH

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

REM2 HUMHI 80 %RH

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------



Press DOWN to decrease the appropriate setpoint. Maintain the switch pressed to make larger changes.

Press UP to increase the appropriate setpoint. Maintain the switch pressed to make larger changes.

Press SET when the new setting is correct. This instates the value just entered using the UP/DOWN keys as the new alarm setpoint. The menu reverts to the previous screen allowing another setpoint to be altered

Press RESET to revert to the original value prior to pressing UP or DOWN.

Press EXIT to escape back to the previous screen



4.7 CALIBRATION MENU

WARNING

This section of the manual merely describes the layout of the calibration menus. Please refer to Section 6 for details of how to perform calibration.

DO NOT PRESS THESE BUTTONS UNLESS YOU ARE FAMILIAR WITH THE INSTRUMENT. ALTHOUGH THE SYSTEM PROTECTS AGAINST SOME FORMS OF ACCIDENTAL CALIBRATION, IT IS STILL POSSIBLE TO AFFECT THE ACCURACY OF GAS/PRESSURE MEASUREMENT.

From the Main Menu, pressing 'CALIB' brings up a new menu requesting the user to either select REM1, REM2, REM3 or INTERFACE (to set up the internal interface card providing relay and analog outputs).

Calibration

REM1	REM2	REM3	IFACE	MAIN
------	------	------	-------	------

Press REM1 to select the REM1 unit (O2, CO2 and Pressure)

Press REM2 to select the REM2 unit (Temperature and Humidity)

Press REM3 to select the REM3 unit (O2 injection)

Press IFACE to select the relay and analog output interface module - refer Section 4.8

Press 'MAIN' to return back to the main menu.

Having Pressed CALIB/REM1 the following menu appears.

Calibrate REM1

O2	CO2	PR	EXIT	MAIN
----	-----	----	------	------

Press 'O2' to select the options allowing calibration of the Oxygen sensor.

Press 'CO2' to select the options allowing calibration of the Carbon Dioxide sensor.

Press 'PR' to select the options allowing calibration of the Pressure (Depth) sensor.

Press 'EXIT' to return back to the previous menu



Press 'MAIN' to return back to the main menu.



Calibrate REM1 O2

CAL-L	CAL-H		EXIT	MAIN
-------	-------	--	------	------

Calibrate REM1 CO2

ZERO	SPAN		EXIT	MAIN
------	------	--	------	------

Calibrate REM1 PR

CAL-L	CAL-H		EXIT	MAIN
-------	-------	--	------	------

Press CAL-L to calibrate the appropriate sensor at a low point in its output.

For oxygen this could typically be 0% Oxygen or air (20.9% Oxygen). (Note the sensor must have been subjected to the correct calibration gas prior to this procedure, and the reading must have settled)

Press CAL-H to calibrate the appropriate sensor at a high point in its output.

For oxygen this could typically be 100% Oxygen. (Note the sensor must have been subjected to the correct calibration gas prior to this procedure, and the reading must have settled). Also note that the partial pressure of oxygen must not exceed the range of the instrument. Therefore when using 100% oxygen gas, the pressure reading must be less than 2 BarA for standard systems (or 1.1BarA for Black Sky system).

Similarly for REM2 (CALIB/REM2)

Calibrate REM2

TMP	HUM		EXIT	MAIN
-----	-----	--	------	------

Press 'TMP' to select the options allowing calibration of the Temperature sensor.

Press 'HUM' to select the options allowing calibration of the Humidity sensor.

Press 'EXIT' to return back to the previous menu

Press 'MAIN' to return back to the main menu.

Calibrate REM2 TMP

CAL-L	CAL-H		EXIT	MAIN
-------	-------	--	------	------

Calibrate REM2 HUM

CAL-L	CAL-H		EXIT	MAIN
-------	-------	--	------	------

And similarly for REM3 (CALIB/REM3)

Calibrate REM3

O2			EXIT	MAIN
----	--	--	------	------



Press 'O2' to select the options allowing calibration of the Oxygen sensor.
Press 'EXIT' to return back to the previous menu
Press 'MAIN' to return back to the main menu.



Calibrate REM1 O2

CAL-L

CAL-H

EXIT

MAIN

For any of the above REM1, REM2 or REM3 options, selecting one of the calibration options (CAL-L, CAL_H (or SPAN for CO2)) brings up a screen requesting the user to define the calibration parameters. (Note there is no corresponding item for CO2 Zero, as this is always performed using zero CO2).

For oxygen CAL-L (REM1 or REM3)

Cal Data O2 CAL-L 20.9 %

DOWN

UP

SET

RESET

EXIT

O2 CAL-L can be varied in the range 0.0% to 30.0%. This is the percentage content as specified on the calibration gas bottle. The system automatically allows for the pressure at the time of calibration.

For oxygen CAL-H (REM1 or REM3)

Cal Data O2 CAL-H 100.0 %

DOWN

UP

SET

RESET

EXIT

O2 CAL-H can be varied in the range 50.0 to 100.0%. This is the percentage content as specified on the calibration gas bottle. The system automatically allows for the pressure at the time of calibration.

For CO2 Zero

Press ZERO to calibrate the appropriate carbon dioxide sensor in zero concentration CO2 (pure oxygen or pure nitrogen or a mixture of oxygen and nitrogen certified free from CO2). There is no data screen that comes up to allow the user to vary the gas content - it must be Zero CO2.

For CO2 Span,

Cal Data CO2 CAL-H 0.50%

DOWN

UP

SET

RESET

EXIT

CO2 Span can be varied in the range 0.50 - 5.00% (may change with range of sensor). This is the percentage content as specified on the calibration gas bottle. The system automatically allows for the pressure at the time of calibration.

Note that the partial pressure of carbon dioxide must not exceed the range of the instrument. Therefore when using 0.5% gas, the pressure reading must be less than 10 Bar Absolute. It is recommended to perform calibration at atmospheric pressure.



For Pressure CAL-L

Cal Data PR CAL-L 1.00 BarAbs

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

PR CAL-L can be varied in the range 0.00 to 2.00 Bar Abs. Note the pressure must be defined in Bar Absolute, as this is the internal working range of the instrument. For alternative units, the conversion factors applied are 10MSW=1BarA, 32.808FSW = 1 BarA, 1.01325 BarAbs =1.0 ATA

For Pressure CAL-H

Cal Data PR CAL-H 2.00 BarAbs

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

PR CAL-H can be varied in the range 0.95 to 10.00 BarAbs (this will change with range of instrument). Note the pressure must be defined in Bar Absolute as above for CAL-L.

For Temperature CAL-L

Cal Data TMP CAL-L 0.0 °C

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

TMP CAL-L can be varied in the range 0.0 to 25.0 °C

For Temperature CAL-H

Cal Data TMP CAL-H 62.0 °C

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

TMP CAL-H can be varied in the range 30.0 to 50.0 °C

For Humidity CAL-L

Cal Data HUM CAL-L 20.9 %

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------

HUM CAL-L can be varied in the range 0.0% to 40.0%

For Humidity CAL-H

Cal Data HUM CAL-H 100.0 %

DOWN	UP	SET	RESET	EXIT
------	----	-----	-------	------



HUM CAL-H can be varied in the range 50.0 to 100.0%

For any of these screens

Press DOWN to decrease the appropriate setpoint. Maintain the switch pressed to make larger changes.

Press UP to increase the appropriate setpoint. Maintain the switch pressed to make larger changes.

Press SET when the new setting is correct. This instates the value just entered using the UP/DOWN keys as the new calibration data. The menu reverts to the previous screen allowing another data value to be altered

Press RESET to revert to the original value prior to pressing UP or DOWN.

Press EXIT to escape from the calibration sequence.

NOTE : When pressing SET, the calibration data value is transmitted to the Remote Sensor. The Remote Sensor must be switched on whilst performing these tasks.

After defining the calibration value for any of the parameters above, a confirmation screen will appear.

For instance, for REM1 Oxygen CAL-L:

Calibrate REM1 O2 CAL-L



To continue the calibration, the user must press the YES option. Pressing NO or MAIN will not perform the calibration. The sensor must be subjected to the calibration conditions at the time of pressing YES.

After confirming the calibration, the Main Panel sends a command to the Remote Sensor instructing it to calibrate. The user should observe the reading on the display change to the new value. (Note that for CO2 zero, this change takes place over a few seconds).

If the Main Panel determines that the sensor reading is invalid, a calibration alarm is raised (CAL-L or CAL-H). This must be cleared by either repeating the calibration sequence correctly, or repairing a fault causing the problem.

4.8 INTERFACE CALIBRATION MENU

From the Calibration Menu, pressing 'IFACE' brings up a screen that allows calibration of the relay and analog output interface optionally fitted internally within the Main Panel.

Note these options will only operate correctly if the 'RLY' indicator on the top line of the display is NOT showing.

Calibrate/Test Interface

ANOUT RELAY MAIN

Press ANOUT to select the Analog Output option

Calibrate Analog :

ZERO MID SPAN OVER MAIN

Press ZERO to force each of the outputs to Zero (4mA for a 4-20mA output)

Press MID to force each of the outputs to mid-scale (12mA for a 4-20mA output)

Press SPAN to force each of the outputs to full-scale (20mA for a 4-20mA output)

Press OVER to force each of the outputs to overscale (fault) condition (approx 28-30mA for a 4-20mA output)

Press MAIN to exit, and to restore the analog outputs to their normal (ie measured value) indications.

Press CALIB/IFACE/RELAY to select the Relay Output option

Test Relay Outputs: 00000000

NEXT OFF ON MAIN

The Binary byte shows the status of each of the relays within the system. A '0' indicates that the relay is de-energised, a '1' indicates that the relay is energised.

Bit Position	7	6	5	4	3	2	1	0
Relay	N/A	N/A	1	2	3	4	5	6

Press NEXT to advance a reverse highlight cursor to the bit to be altered.

Press OFF to clear the highlighted bit

Press ON to set the highlighted bit.



4.9 SETUP MENU

From the Main Menu, press 'SETUP' . The Setup Menu appears

Setup Menu

TIME	LOG			MAIN
------	-----	--	--	------

Press 'TIME' to adjust the date and time

Press 'LOG' to alter the data logging time interval.

Press 'MAIN' to return to the Main Menu.

12:34:56 01-Feb-2000

DOWN	UP	NEXT	SET	MAIN
------	----	------	-----	------

Initially the hours field will be highlighted. Press UP or DOWN to alter the hours.

Then press NEXT to highlight the minutes field. Press UP or DOWN to alter the minutes.

Then press NEXT to select the Seconds, Day, Month and Year fields as required.

When no further changes are required, press SET. If an invalid date has been selected, an 'Invalid' message will appear. Correct the date or press MAIN to abort entry of the date and time.

Select Data Log Interval

NEXT				MAIN
------	--	--	--	------

Press NEXT to select the Data Logging time interval as either 1 second, 5 seconds, 10 seconds, 30 seconds, 1 minute, 2 minutes or 5 minutes. This option can be altered at any time, irrespective of whether data logging is in progress.



4.10 OXYGEN INJECTION MENU

From the Main Menu, pressing O2 INJCT brings up the injection menu (on systems with this option fitted)

Oxygen Injection Control



Press OFF to switch off Injection control - the status on the Injection screen will change to OFF.

Press ON to turn on Injection control - the status on the Injection screen will change to ON, unless the switch on the REM3 sensor is in the DISABLE position, in which case the display will show 'Dis'.

Press SETPT to adjust the oxygen injection setpoint

Press MIX to adjust the oxygen injection Mixing time

Press MAIN to return to the Main Menu.

Pressing SETPT, brings up a screen to alter the injection setpoint

O2 Inj Setpoint = 209 mBar



Press DOWN to decrease the setpoint. Maintain the switch pressed to make larger changes.

Press UP to increase the setpoint. Maintain the switch pressed to make larger changes.

Press SET when the new setting is correct. This instates the value just entered using the UP/DOWN keys as the new setpoint, and transmits the new setting to the REM3. The menu reverts to the previous screen allowing another option to be selected

Press RESET to revert to the original value prior to pressing UP or DOWN.

Press EXIT to escape back to the previous screen

Pressing MIX, brings up a screen to alter the injection mixing time

O2 Inj Mixing Time = 60 sec



Press DOWN to decrease the setpoint. Maintain the switch pressed to make larger changes.

Press UP to increase the setpoint. Maintain the switch pressed to make larger changes.

Press SET when the new setting is correct. This instates the value just entered using the UP/DOWN keys as the new setpoint, and transmits the new setting to the REM3. The menu reverts to the previous screen allowing another option to be selected

Press RESET to revert to the original value prior to pressing UP or DOWN.

Press EXIT to escape back to the previous screen

5 DATALOGGING

A 25 way Female D-type connector is fitted to the rear of the Main Panel This is intended for Diagnostic purposes and for retrieving data logged information from the system.

The Diagnostic Connector details are shown below

Pin Number	Signal	Purpose
2	RS232 Transmit Output	Data Communications with an External computer
3	RS232 Receive Input	
7	RS232 Signal Ground	
9	24V DC Output	Power Supply output for Test Purposes only
11	0V DC Output	
13	RS485 A-	RS485 connections for listening to communications (Factory Use Only)
25	RS485 B+	

5.1 RS232 DATA LOGGING CONNECTION

An RS232 connection must be made between a computer and the Diagnostic Port to use this facility.

A typical lead to connect a standard IBM PC compatible computer/laptop would consist of either a 9 or a 25 way D-type female connector connected to a 25 way male D-type connector. A 9 way connector lead is supplied with the system. Refer to the diagram below for alternative connections.

DIAGNOSTIC PORT 25 WAY MALE D-TYPE	PC SERIAL PORT	
	9 WAY FEMALE D-TYPE	25 WAY FEMALE D-TYPE
Pin 2	Pin 2	Pin 3
Pin 3	Pin 3	Pin 2
Pin 7	Pin 5	Pin 7



At power-up the data logging will always default to OFF.

To commence logging, press LOG TOGGLE on the Main Menu. Assuming the memory is not full, the Logging Status will change to ON.

During logging, the value of each measured parameter (Oxygen, CO₂ and Pressure and Temperature and Humidity if a REM2 is fitted) will be recorded every 60 seconds. This interval may be changed to an interval between 1 second and 5 minutes, as described in the Function Key section of this manual.

Data logging is stopped by pressing LOG TOGGLE again.

Every logged reading is stored with the date and time. Successive log runs can therefore be stored consecutively without having to delete the previous readings.

The state of the Data Log Buffer is indicated on the screen as %FULL. Since the unit can store several thousand readings, this will remain at a low number for a considerable time.



5.2 USE OF WINDOWS HYPERTERMINAL

A suitable communications programme must be run on the computer to access the serial port. HYPERTERM.EXE provided with Windows (95, 98, 2000, XP) will suffice, or several common terminal programmes such as Procomm or Kermit.

The Communication parameters required for successful communication are 9600 Baud, 8 Data Bits, No Parity, 1 Stop Bit, XON/XOFF handshaking. Make sure the identity of the COM port on the PC is known (COM1, COM2... etc). If not clear, this information can be found with Control Panel, System, Hardware, Device Manager, Ports.

1	Ensure that HyperTerminal has been installed. It is supplied by Microsoft with an original Windows installation. If it is not installed, go to Control Panel, Add/Remove Programs, Windows Setup, and select HyperTerminal from the Communications Utilities.
2	Start HyperTerminal. Usually this will be found under Start, Programs, Accessories, Communications, HyperTerminal, HyperTerm.exe
3	Enter a name for the connection (eg Analox) and click OK.
4	Select 'Connect Using', typically COM 1 or COM 2.
5	Update parameters as follows: Baud - 9600 Data Bits - 8 Parity - None Stop Bit - 1 Flow Control - XON/XOFF
6	Click 'OK'
7	Switch the Analox Sub On
8	If the software is correctly configured and the correct connections made, switching the system on will result in a message 'dd-mmm-yyyy hh:mm:ss System Started '. The dd-mmm-yyyy and hh:mm:ss indicates the date and time on the microprocessor card.
9	If no text appears, check the connections and the communication parameters. Also check that there is no conflict between the selected COM port and any other devices already in use on the PC.
10	Typing 'HELP' followed by pressing Enter will bring up a Help menu on the PC screen as shown below. The case of 'help' is not important - upper or lower case can be used.

ANALOX SUB MK2F AMS-310-xx dd-mmm-yyyy



```

HELP => Display this message
GET  => Read Data Logged Information
DEL  => Delete Data Logged Information
EXIT => Shutdown

```

All of the commands must be followed by pressing Enter. Use of these commands is discussed in the next section.

5.3 RETRIEVE DATA: GET COMMAND

To download data from the instrument, first set the PC HYPERTRM.EXE program to capture data to file. This is done by selecting the option Transfers, Receive Text File.. Enter a suitable name for a file in which to store data, and note in which directory it is defined. Then press OK.

Note that data logging must be OFF for this to operate. If necessary press LOG TOGGLE to turn logging OFF

Now type GET followed by ENTER.

Data received will be in the form

```

DATE,TIME,OXYGEN,CO2,DEPTH,TEMPERATURE,HUMIDITY,INJECTION
10-Jul-2000 08:47:10, 207, 0.3, 0.99, 20.0, 43, 209
10-Jul-2000 08:47:20, 207, 0.4, 0.99, 20.0, 43, 209
Data Output Completed

```

One line of data will appear for every sample of readings logged. The system can store several thousand readings. Retrieval of data may therefore take a while to complete.

When the 'Data Output Completed' message appears, press the STOP button on the PC screen to stop recording to file.

5.4 DELETE DATA: DEL COMMAND

The DEL command is used to delete the stored information from the system. Perform this command after successful retrieval of the data using GET. If DEL is not used, the same data will have to be retrieved again at the next GET command. For DEL to operate, data logging must be OFF. Press LOG TOGGLE if necessary.



5.5 USE OF RETRIEVED DATA

It is assumed that the user has already created a file in a known directory by performing the GET command. This file can be imported into any of the major spreadsheet (Excel, Lotus 123 etc) since it is stored in Comma Separated format.

Once the data has been imported into a spreadsheet, the data can be graphed and printed. Refer to the manual for your preferred spreadsheet.

Scottish Anglo can provide detail for an example data file using Microsoft Excel if required. The appearance of the graphs generated is then only dependant on the users ability to manipulate the spreadsheet.

6 CALIBRATION

Calibration should only be performed by personnel with the necessary skills. Any abuse of the calibration controls may render the instrument inaccurate and unusable.

For calibrating the oxygen and carbon dioxide sensors, the supplied calibration adaptors and tubing should be used as shown in the drawing below. Remove the adaptors from the sensors after calibration. The drawing also shows a calibration gas bottle and flow regulator. These can be purchased from Analox if necessary. Alternatively the user may opt to use their own calibration gas, perhaps with a conventional pressure regulator and an in-line sample flow meter. Take care not to over-pressurise the sensors, or they may be damaged.




Figure 9: Calibration Arrangement

6.1 SEMI AUTOMATIC GAS CALIBRATION

The system features a semi-automatic calibration feature for zero and span of oxygen, carbon dioxide and pressure.

These adjustments are possible without internal access to the sensor unit, provided that the sensors are near to their ideal outputs.

The depth sensor is used in measuring the carbon dioxide content, therefore before altering the carbon dioxide calibration, ensure that the depth sensor is correctly calibrated.

The depth sensor is also used in measuring the oxygen content if oxygen is measured as a percentage.

Calibration of both oxygen and carbon dioxide requires that the depth sensor is correctly calibrated..

To ensure the most accurate performance, ensure that settling time is allowed for all gas readings. Typically 2-5 minutes is required for adequate settling time.

6.2 PRESSURE SENSOR CALIBRATION

1	Subject the remote sensor to a known pressure - say 1.00BarAbsolute
2	Ensure that the reading on the display is steady
3	From Main menu, press CALIB, REM1, PR, CAL-L and adjust the setting to the actual pressure in Bar Absolute (eg 1.000BarA).
4	Press SET to inform the remote sensor of the calibration pressure
5	Now confirm the action by pressing YES
6	Observe after a few seconds that the pressure reading on the display adjusts to the new calibration value
7	Now subject the remote sensor to a known pressure - say 9.00 Bar Absolute
8	Ensure that the reading on the display is steady
9	From Main menu, press CALIB, REM1, PR, CAL-H and adjust the setting to the actual pressure in Bar Absolute. (eg 8.980 BarA)
10	Press SET to inform the remote sensor of the calibration pressure
11	Now confirm the action by pressing YES
12	Observe after a few seconds that the pressure reading on the display adjusts to the new calibration value.
13	Reduce the pressure to a mid scale value and confirm that the reading on the instrument is correct.

6.3 OXYGEN SENSOR CALIBRATION

1	Fit the supplied calibration adaptors to the sensor inlets and pass calibration gas of a certified concentration (typically calibrated air 20.9% or nitrogen 0%) across the sensors at a flow rate of approximately 20-60 litres per hour (0.3-1.0 litres per minute)
2	Wait for the sensor reading to stabilise
3	From the MAIN menu press CALIB, REM1, O2, CAL-L and adjust the gas concentration to the desired value (eg 0.0 for nitrogen, or 20.9% for air etc) The actual gas can be in the range 0-30.0%.
4	Press SET to inform the remote sensor of the calibration gas value
5	Now confirm the action by pressing YES
6	Observe after a few seconds that the oxygen reading adjusts to the new calibration value (taking the absolute pressure into account)
7	Now pass a higher concentration certified calibration gas (typically 100% O2) across the sensors at a flow rate of approximately 20-60 litres per hour (0.3-1.0 litres per minute)
8	Wait for the sensor reading to stabilise
9	From the MAIN menu press CALIB, REM1, O2, CAL-H and adjust the gas concentration to the desired value (eg 100.0 %) The actual gas can be in the range 50-100% Press SET to inform the remote sensor of the calibration gas value.



10	Now confirm the action by pressing YES
11	Observe after a few seconds that the oxygen reading adjusts to the new calibration value (taking the absolute pressure into account)
12	For systems with a REM3 Injection unit, repeat the above process using CALIB, REM3 options on the Main Panel
13	Remove the calibration flow adaptors after use and confirm that the sensor readings are sensible for the atmosphere in which they are exposed.

6.4 CARBON DIOXIDE SENSOR CALIBRATION

1	Fit the supplied calibration adaptors to the sensor inlets and pass calibration gas containing no carbon dioxide (typically calibrated air 20.9% or nitrogen 0% or 100% oxygen) across the sensors at a flow rate of approximately 20-60 litres per hour (0.3-1.0 litres per minute)
2	Wait for the sensor reading to stabilise
3	From the Main menu press CALIB, REM1, CO2, ZERO
4	Confirm calibration by pressing YES.
5	Observe after a few seconds that the carbon dioxide reading adjusts to zero.
6	Now pass a certified calibration gas (typically 2.0% CO2) across the sensors at a flow rate of approximately 20-60 litres per hour (0.3-1.0 litres per minute)
7	Wait for the sensor reading to stabilise
8	From the MAIN menu press CALIB, REM1, CO2, SPAN and adjust the gas concentration to the desired value (eg 2.05%) The actual gas can be in the range 0.50-5.00% (depending on range)
9	Press SET to inform the remote sensor of the calibration gas value.
10	Now confirm the action by pressing YES
11	Observe after a few seconds that the carbon dioxide reading adjusts to the new calibration value (taking the pressure into account).
12	Remove the calibration flow adaptors after use and confirm that the sensor readings are sensible for the atmosphere in which they are exposed.

6.5 TEMPERATURE SENSOR CALIBRATION

1	Subject the REM2 Sensor unit to a known temperature in the range 0-20° C, and allow the reading to stabilise..
2	From the Main menu press CALIB, REM2, TEMP, CAL-L
3	Enter the actual temperature to which the REM2 has been subjected. Press SET after obtaining the correct temperature value.
4	Now confirm the action by pressing YES
5	Observe after a few seconds that the temperature reading adjusts to the correct value.



6	Now subject the REM2 sensor to a higher temperature (20-50°C), and allow the reading to stabilise..
7	From the Main menu press CALIB, REM2,TEMP,CAL-H and enter the actual temperature to which the REM2 has been subjected. Press SET after obtaining the correct temperature value.
8	Now confirm the action by pressing YES
9	Observe after a few seconds that the temperature reading adjusts to the correct value.

6.6 HUMIDITY SENSOR CALIBRATION

1	Subject the REM2 Sensor unit to a known humidity in the range 0-40%RH.
2	Allow the reading to stabilise.
3	From the Main menu press CALIB, REM2, HUM, CAL-L and enter the actual humidity to which the REM2 has been subjected.
4	Press SET after obtaining the correct humidity value.
5	Now confirm the action by pressing YES
6	Observe after a few seconds that the humidity reading adjusts to the correct value.
7	Now subject the REM2 sensor to a higher humidity (50-100%RH)
8	Allow the reading to stabilise.
9	From the Main menu press CALIB, REM2,HUM,CAL-H and enter the actual humidity to which the REM2 has been subjected.
10	Press SET after obtaining the correct humidity value.
11	Now confirm the action by pressing YES
12	Observe after a few seconds that the humidity reading adjusts to the correct value.

6.7 INTERNAL CALIBRATION

Note: in normal operation, potentiometers should only be adjusted by suitably qualified personnel. ANY ABUSE OF THE CALIBRATION CONTROLS WILL ADVERSELY AFFECT THE ACCURACY OF THE INSTRUMENT AND MAY RENDER THE INSTRUMENT UNUSABLE. AFTER MAKING ANY CHANGES TO THE POTENTIOMETERS, THE SOFTWARE CALIBRATION PROCEDURES SHOULD ALWAYS BE PERFORMED. The positions of potentiometers referenced in the following sections are shown in Figure 10 and Figure 11.

6.7.1 Oxygen sensor calibration

1	Pass nitrogen over the oxygen sensor (top sensor) at a flow rate of approximately 20-60 litres per hour
2	Whilst supporting the enclosure lid, undo the four screws at the corners of the REM1 enclosure.
3	If the oxygen reading is greater than zero, make small anti-clockwise adjustments to the O₂ ZERO potentiometer (RV4) until the reading JUST changes to zero. If the oxygen reading is zero, make small clockwise adjustments to the O₂ ZERO potentiometer (RV4) until the reading just rises above zero, then make a small anti-clockwise adjustment so that the reading JUST changes to zero.
4	Remove the flow of nitrogen and pass a known concentration of oxygen over the sensor at a flow rate of approximately 20-60 litres per hour. The oxygen concentration should ideally be around 50-90% of full scale for accurate calibration.
5	The O ₂ SPAN potentiometer (RV5) is not intended to be adjusted. If the software calibration feature proves inadequate, RV5 may be adjusted until the correct concentration of oxygen is displayed (clockwise adjustment increases the reading, anti-clockwise decreases).

6.7.2 CO₂ sensor calibration

Calibration of the CO₂ sensor can only be performed by suitably qualified personnel with the correct diagnostic equipment and therefore should not be attempted. Please contact your service agent if a CO₂ sensor calibration is required.





6.7.3 Pressure sensor calibration

Re-calibration of the pressure sensor requires the adjustment of the pressure ZERO potentiometer (RV6) at atmospheric pressure, and adjustment of the pressure SPAN potentiometer (RV7) at an increased pressure - preferably around 90% of full scale. SPAN adjustment can be performed at a pressure lower than 90% of full scale, but this will result in a slightly less accurate calibration.

1	Whilst supporting the enclosure lid, undo the four screws at the corners of the REM1 enclosure.
2	With the sensor subjected to atmospheric pressure only, adjust pot. RV6 until the current atmospheric pressure is displayed on the REM1 pressure display.
3	Pressurize the sensor to approximately 90% of full scale and adjust pot. RV7 until the correct pressure is displayed on the REM1 display.
4	Repeat steps 2 and 3 until no further potentiometer adjustment is necessary. Replace the enclosure lid.

Note that systems rated up to 10 Bar Absolute use a small internal pressure sensor. It is possible to connect a tube to the pressure port of this device, but it is difficult to keep it clamped on at pressures above around 5 BarA. Doing so, avoids the need to pressurise the whole instrument. Take extreme care not to exert any undue forces on the pressure port as it is easily damaged.

For systems rated at greater than 10BarA, a pressure sensor is mounted below the oxygen and CO₂ sensors. The pressure port of this sensor is a male G1/4. It is possible to make a connection to this port to connect the pressure sensor to a pressure calibrator.

6.7.4 Temperature sensor calibration

1	Whilst supporting the enclosure lid, undo the four screws at the corners of the REM2 enclosure
2	Remove the connector from JP6 (Temperature Sensor).
3	Connect a milli-ammeter across the pins of JP6.
4	Adjust RV1 to achieve a current reading of 1.00mA.
5	Reconnect JP6
6	Now connect a voltmeter across TP9 (+) and TP8 (-).
7	Ensure SW2 is in the '0' position.
8	Maintain the RUN/CAL switch to CAL.
9	Adjust RV7 to achieve a voltage reading of approximately 0.2v.
10	Ensure SW2 is in the '39' position.
11	Maintain the RUN/CAL switch to CAL.



12	Adjust RV3 to achieve a voltage reading of approximately 2.0v
13	Repeat from Step 7 until no further adjustment is necessary
14	Perform the software calibration to complete the calibration.

Nb Older versions of instruments were span calibrated at 62 °C rather than 39 °C. The corresponding voltage for 62 °C was 2.5V.




6.7.5 Humidity sensor calibration

1	Whilst supporting the enclosure lid, undo the four screws at the corners of the REM2 enclosure
2	Now connect a voltmeter across TP10 (+) and TP8 (-). This signal typically varies between 0.2 and 3.0v for 0-100% humidity.
3	Subject the sensor to a 0% humidity and adjust RV8 to achieve a voltage reading of approximately 0.2v.
4	Subject the sensor to a 100% humidity and adjust RV5 to achieve a voltage reading of approximately 3.0v.
5	Repeat from Step 4 until no further adjustment is necessary
6	Perform the software calibration to complete the calibration.

6.7.6 Analogue Output Calibration

The analogue outputs are factory set to provide 4-20mA outputs representing 0 to full scale for each channel. Voltage outputs may also be provided by prior arrangement at the time of manufacturing.

1	Remove the four screws fastening the rear panel of the Main Panel unit - this is best done on a suitable flat surface.
2	Withdraw the rear panel assembly out of the enclosure
3	Identify the ASB-251 printed circuit board mounted in the base of the unit
4	Identify the potentiometers for each of the analogue output channels as shown in Figure 12..
5	From the Main Menu, press CALIB, IFACE, ANOUT.
6	Press ZERO
7	Adjust the Zero potentiometer to achieve 4.00mA in the output circuit.
8	Press SPAN
9	Adjust the Span potentiometer to achieve 20.00mA in the output circuit
10	Repeat from pressing ZERO until no further adjustments of either ZERO or SPAN are required
11	Press MID and confirm the output current in 12.0mA
12	Then repeat the sequence for each channel.
13	Press MAIN to restore all of the outputs to their normal values.
14	When all channels are correctly adjusted, refit the rear panel to the enclosure and fasten the 4 screws.



ASB-541-02

Figure 10: REM1 Calibration pot positions




Figure 11: REM2 Calibration pot positions




Figure 12: Interface Board - Calibration Pot Positions

7 TROUBLESHOOTING

SYMPTOM	REASON	SOLUTION
Main panel does not switch on (Note it takes approx 40s for the system to switch on due to the time taken for the internal processor to boot)	No power connected Rear panel mains fuse blown Front panel DC fuse F1 blown	Ensure power is connected correctly Switch it to ON Replace fuse if necessary. If replacement blows, seek assistance
Remote unit does not switch on	Main unit is switched OFF Fuse F2 on main unit blown Wiring incorrect	Switch main unit ON Replace fuse. If replacement blows, seek assistance Check wiring correct
Fault Indicator is ON and main unit sensor readings are replaced by 'COMMS ERROR' message.	Lack of communications with remote unit	Check remote has power and that wiring agrees with Installation drawings.
Fault Message appears on Display		Please refer to Section 4.3.2.

8 SAFETY WARNINGS

The oxygen sensor is an electrochemical device and contains a caustic electrolyte. Always check to make sure that it is not leaking and do not allow it onto any part of your body or clothing.

When the life of the sensor has expired or it is leaking or otherwise damaged it must be disposed of safely in accordance with local regulations.

The sensor contains Potassium Hydroxide solution (KOH) which is hazardous and can have the following effects:

Body Part	Effect	First Aid Procedures
Skin	Contact could result in a chemical burn.	Wash the affected part with a lot of water and remove contaminated clothing. If stinging persists get medical attention.
Ingestion	Can be harmful or FATAL if swallowed.	Drink a lot of fresh water. Do not induce vomiting Get medical help immediately
Eye	Contact can result in the permanent loss of sight.	Wash with a lot of water for at least 15 minutes and get medical help immediately



9 SPECIFICATION

Power Source	External Stabilised 12-24V DC supply with regulation of better than +/- 300mV. External 110-230V AC supply (Option only) The nature of the supply must be specified at the time of ordering.	
Operating Current	Average 400mA at 24V DC supply, peak current 550mA Power supply must be capable of providing instantaneous peak 2A at switch on	
Fuses	F1 1A-M in supply to Main Panel F2 400mA Fast Acting in supply to REM1 Both fuses are mounted on the front panel Mains powered instruments also have a 500mA-T fuse on the rear panel	
Display Panel	LCD graphic display, 240 x 128 pixels, with adjustable backlight and contrast controls Displays time in hours:minutes:seconds in 5mm high characters Displays current values of O ₂ , CO ₂ , Pressure, Temperature and Humidity using 10mm high large characters	
Alarm Indicators	1 Alarm indicator for gas/environment alarms 1 Fault indicator for communications, calibration faults. 1 Audible Buzzer operating on alarm/fault conditions.	
Operator controls	1 Power Switch to switch instrument on and off 5 Pushbuttons used via Menu System 1 Pushbutton to mute alarms	
Oxygen Sensors	Analox 9100-9212-9HSUB oxygen sensor with up to 3 year life at 0.21 ATA ppO ₂ . Displayed as mBar, %SEV, Percent or ATA (define at time of order)	
	Ranges	0-2000 mBar
	Accuracy	< 1% of range at constant temperature and pressure
Pressure Sensor	Analox pressure transducer, with bridge output. Displayed as BarA, BarG, MSW, FSW or ATA(define at time of order)	
	Range	0-10.00 BarA or 0-60.00 BarA (other ranges by request)
	Accuracy	< 1% of range
CO ₂ Sensor	Analox BL5 low power, long life infra red sensor. CO ₂ reading is pressure compensated by microprocessor Displayed as mBar, %SEV, Percent or ATA (define at time of order)	

	Accuracy	<5% of range for 0-midscale and for 0.8BarA<Pressure <6BarA <10% of range for midscale-fullscale and for Pressure <0.8BarA or >6BarA
	Range	0-20mBar, 0-50mBar, 0-100mBar (define at time of order)
Analogue Outputs	Optional Up to 5x 4-20mA output channels - (active outputs - generating current output) Voltage outputs also available by request.	
Relay Outputs	Optional Up to 6x Relay outputs assignable to any alarm or fault conditions. Changeover contacts on each relay, normally configured in Fail-Safe Mode. Contacts rated at 0.5A (230V AC) or 1 Amp (30 VDC)	
Operating Temperature	0°C to 40°C	
Storage Temperature	-30°C to 55°C (remove oxygen sensor below -10°C)	
Dimensions	Main Panel	240mm (w) x 133mm (h) x 252mm (d)
	REM1/2	160mm (w) x 120mm (h) x 90mm (d)
Weight	Main Panel	3.0kg
	REM1	0.75kg
	REM2	0.75kg

По вопросам продаж и поддержки обращайтесь:

Алматы (727)345-47-04
Ангарск (3955)60-70-56
Архангельск (8182)63-90-72
Астрахань (8512)99-46-04
Барнаул (3852)73-04-60
Белгород (4722)40-23-64
Благовещенск (4162)22-76-07
Брянск (4832)59-03-52
Владивосток (423)249-28-31
Владикавказ (8672)28-90-48
Владимир (4922)49-43-18
Волгоград (844)278-03-48
Вологда (8172)26-41-59
Воронеж (473)204-51-73
Екатеринбург (343)384-55-89

Иваново (4932)77-34-06
Ижевск (3412)26-03-58
Иркутск (395)279-98-46
Казань (843)206-01-48
Калининград (4012)72-03-81
Калуга (4842)92-23-67
Кемерово (3842)65-04-62
Киров (8332)68-02-04
Коломна (4966)23-41-49
Кострома (4942)77-07-48
Краснодар (861)203-40-90
Красноярск (391)204-63-61
Курск (4712)77-13-04
Курган (3522)50-90-47
Липецк (4742)52-20-81

Магнитогорск (3519)55-03-13
Москва (495)268-04-70
Мурманск (8152)59-64-93
Набережные Челны (8552)20-53-41
Нижний Новгород (831)429-08-12
Новокузнецк (3843)20-46-81
Ноябрьск (3496)41-32-12
Новосибирск (383)227-86-73
Омск (3812)21-46-40
Орел (4862)44-53-42
Оренбург (3532)37-68-04
Пенза (8412)22-31-16
Петрозаводск (8142)55-98-37
Псков (8112)59-10-37
Пермь (342)205-81-47

Ростов-на-Дону (863)308-18-15
Рязань (4912)46-61-64
Самара (846)206-03-16
Санкт-Петербург (812)309-46-40
Саратов (845)249-38-78
Севастополь (8692)22-31-93
Саранск (8342)22-96-24
Симферополь (3652)67-13-56
Смоленск (4812)29-41-54
Сочи (862)225-72-31
Ставрополь (8652)20-65-13
Сургут (3462)77-98-35
Сыктывкар (8212)25-95-17
Тамбов (4752)50-40-97
Тверь (4822)63-31-35

Тольятти (8482)63-91-07
Томск (3822)98-41-53
Тула (4872)33-79-87
Тюмень (3452)66-21-18
Ульяновск (8422)24-23-59
Улан-Удэ (3012)59-97-51
Уфа (347)229-48-12
Хабаровск (4212)92-98-04
Чебоксары (8352)28-53-07
Челябинск (351)202-03-61
Череповец (8202)49-02-64
Чита (3022)38-34-83
Якутск (4112)23-90-97
Ярославль (4852)69-52-93

Россия +7(495)268-04-70

Казахстан +(727)345-47-04

Беларусь +(375)257-127-884

Узбекистан +998(71)205-18-59

Киргизия +996(312)96-26-47

эл.почта: axq@nt-rt.ru || сайт: <https://analox.nt-rt.ru/>